

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

<b>TO</b>	Cher Snyder, Assistant DEO Engineering & Compliance	<b>LABORATORY NO</b>	<u>1609411</u>
		<b>DATE RECEIVED</b>	<u>04/04/2016</u>
<b>SAMPLE DESCRIBED AS</b>	24 hour sample Canister # 54098	<b>FACILITY ID NO</b>	<u>NA</u>
		<b>REQUESTED BY</b>	<u>Sumner Wilson</u>
<b>SAMPLING LOCATION</b>	Highlands Community Pool Parking Lot	<b>ST NO / PROJECT</b>	<u>NA</u>


---

---

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

Type	Canister
Number	<b><u>54098</u></b>
Pressure (Torr)	740
CO, ppm	< 3
CH <sub>4</sub> , ppm	2
CO <sub>2</sub> , ppm	419
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 4/6/16

Approved By:   
Solomon Teffera, Acting Senior Manager  
Laboratory Services  
909-396-2199

4/6-16

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	<u>1609411</u>
24 hour sample Canister # 54098	<b>REQUESTED BY</b>	<u>Sumner Wilson</u>

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

**QUALITY CONTROL -- End of run control recovery**

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.03	5.28	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	2.19	8.20	PASS
CO <sub>2</sub> , ppmvC	0.3	1.57	2.16	37.61	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	2.15	5.78	PASS
NM/NEOC, ppmvC	0.4	2.03	2.27	11.77	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1064	2.70	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1057	-1.07	PASS
CO <sub>2</sub> , ppmvC	0.3	1022	1045	2.29	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	1044	1051	0.69	PASS
NM/NEOC, ppmvC	0.4	1024	1025	0.17	PASS

**DATE ANALYZED** 04/04/2016  
**REFERENCE NO:** 16QM2AB  
 QM2-101-94

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**

DISTRI  
 INVOIC  
 LAP AU  
LABORA

WO #: 1609411



TO: SCAQMD LAB:  OTHER:  \_\_\_\_\_  
 SOURCE NAME: Southern California Gas Co. I.D. No. \_\_\_\_\_  
 Source Address: 12801 Tampa Ave City: Porter Ranch  
 Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: 91326  
 Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: Sumner Wilson Date: 4/4/16  
 Approved by: Jason Low Office: \_\_\_\_\_ Budget #: 44716  
 REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill   
 Suspected Violation Rule(s) \_\_\_\_\_ Other  \_\_\_\_\_

Sample Collected by: Qian Zhou Date: 4/4/16 Time: 9:20am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54098	4-3-16 / 00:00 / 24 hours	-30"	+11.5

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhangjian</i>	<i>Ningning Ran</i>	SCAQMD Lab	4/4/16	11:58

Remarks 1:3 scheduled samples from trailer

Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)