

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1616928</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 53386	DATE RECEIVED	<u>06/18/2016</u>
SAMPLING LOCATION	Castlebay Elementary School	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>53386</u>
Pressure (Torr)	760
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	417
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 6/22/16

Approved By: 
Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

Handwritten notes:
JLW
6-22-16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1616928</u>
24 hour sample Canister 53386	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.60	35.29	PASS
CH ₄ , ppmvC	0.3	2.02	1.99	-1.81	PASS
CO ₂ , ppmvC	0.4	1.57	2.15	36.81	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.10	3.30	PASS
NM/NEOC, ppmvC	0.5	2.03	2.04	0.37	PASS


		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1058	2.11	PASS
CH ₄ , ppmvC	0.3	1068	1055	-1.19	PASS
CO ₂ , ppmvC	0.4	1022	1037	1.45	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1039	-0.48	PASS
NM/NEOC, ppmvC	0.5	1024	1015	-0.84	PASS

DATE ANALYZED	06/18/2016
REFERENCE NO:	16QM2AD QM2-101-112

WO #: 1616928

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT
 INVOICE
 LAP AUD
 LABORATORY NO.



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 6/18/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 6/18/16 Time: 11:10am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem	53386	6/17/16 / 00:00 / 24 hours	-30"	+10.5

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>Ryan J. Bell</i>	SCAQMD Lab	6/18/16	13:31

Remarks: 1:3 scheduled samples from trailer at Castlebay
 Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA
 Left sampler, sn 3419